



U.S. Department
of Transportation
**Pipeline and
Hazardous Materials
Safety Administration**

**IAEA CERTIFICATE OF COMPETENT AUTHORITY
SPECIAL FORM RADIOACTIVE MATERIALS
CERTIFICATE USA/0458/S-96, REVISION 6**

East Building, PHH-23
1200 New Jersey Avenue Southeast
Washington, D.C. 20590

This certifies that the source described has been demonstrated to meet the regulatory requirements for special form radioactive material as prescribed in the regulations of the International Atomic Energy Agency¹ and the United States of America² for the transport of radioactive material.

1. Source Identification - Neutron Products, Inc. Model NPRP 450-10-B.
2. Source Description - Cylindrical double encapsulation made of Type 316L stainless steel with fusion seal welded endcaps. Approximate outer dimensions are 11.1 mm (0.437 in.) in diameter and 451.5 mm (17.777 in.) in length. Construction shall be in accordance with attached Neutron Products, Inc. Drawing No. A-200355, Rev. A, or Neutron Products, Inc. Drawing No. A-200355, Rev. B.
3. Radioactive Contents - Source capsules manufactured in accordance with Neutron Products, Inc. Drawing No. A-200355, Rev. A, shall contain no more than 445 TBq (12,000 Ci) of Cobalt-60. Source capsules manufactured in accordance with Neutron Products, Inc. Drawing No. A-200355, Rev. B, shall contain no more than 500 TBq (13,500 Ci) of Cobalt-60. The Co-60 is in the form of solid metal rods.
4. Quality Assurance - Records of Quality Assurance activities required by Paragraph 310 of the IAEA regulations¹ shall be maintained and made available to the authorized officials for at least three years after the last shipment authorized by this certificate. Consignors in the United States exporting shipments under this certificate shall satisfy the applicable requirements of Subpart H of 10 CFR 71.
5. Expiration Date - This certificate expires on March 31, 2022.

¹ "Regulations for the Safe Transport of Radioactive Material, 1996 Edition (Revised), No. TS-R-1 (ST-1, Revised)," published by the International Atomic Energy Agency(IAEA), Vienna, Austria.

² Title 49, Code of Federal Regulations, Parts 100-199, United States of America.


CERTIFICATE USA/0458/S-96, REVISION 6

This certificate is issued in accordance with paragraph 804 of the IAEA Regulations and Section 173.476 of Title 49 of the Code of Federal Regulations, in response to the February 24, 2017 petition by Neutron Products, Inc., Dickerson, MD, and in consideration of other information on file in this Office.

Certified By:



Mar 03 2017
(DATE)




 _____
William Schoonover
Acting Associate Administrator for Hazardous Materials Safety

Revision 6 - Issued to extend the expiration date.

ENGRAVE
YEAR OF MFG.

NOTES:

1. STAINLESS STEEL TUB AND ENDCAPS 316L
2. LIGHT PRESS FIT INTO TUBING C'BORE
3. SPACE FOR THERMAL EXPANSION
4. ENGRAVING .093 HIGH X .005 - .010 DEEP
5. STAINLESS STEEL SPACERS 300 - SERIES



END CAP DWG A-200031

ENGRAVE

FUSION WELD
END CAPS (4)

INNER ENCAPSULATION

COBALT RODS
.245 DIA X .996 OR .498 Lg.
WITH STAINLESS STEEL
SPACERS AS APPROPRIATE

OUTER ENCAPSULATION
TUBING .382 OD
X .025 WALL
X 16.877 Lg.

ENGRAVE SERIAL No.
SIDE & FACE (UNDERLINED)



3-10-95 0 B 50 L E T E S E E R E F

| | | | | | |
|--|--|--|--|--|--|
| UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES | | NEUTRON PRODUCTS inc | | Dickerson, Maryland | |
| TOLERANCES ANGULAR DECIMAL FRACTIONAL | | TITLE RADIATION PROCESSING SOURCE NPRP 450-10-B | | APPROVED <i>RS</i> | |
| DRAWN ED 11-20-90 | | | | ISSUED DATE 11/21/90 | |
| DESIGN FS 11-20-90 | | | | SIZE DWG. NO. REV. A 200355 A | |
| CHECKED <i>RS</i> | | | | SCALE 2X SHEET 1/21/92 | |

ENGRAVE YEAR
OF MFG.

NOTE:


1. STAINLESS STEEL TUBING AND END CAPS - 316L
2. LIGHT PRESS FIT INTO TUBING C'BORE
3. SPACE FOR THERMAL EXPANSION
4. ENGRAVING .093 HIGH x .005 - .010 DEEP
5. STAINLESS STEEL SPACERS 300 SERIES
6. DIMENSIONS ARE NOMINAL SEE DWG C-200359 (COMPONANTS) FOR TOLERANCES AND ENGRAVING DETAILS



END CAP
INNER ENCAPSULATION

(2 REQ'D)


C'BORE .287 DIA (2 PLCS)



TUBING
INNER ENCAPSULATION

.317 OD x
.025 WALL

17.777 MAX
17.752 MIN



FUSION WELD
END CAPS (4)

INNER ENCAPSULATION

COBALT RODS
.245 DIA x .996 OR .498 LG
WITH STAINLESS STEEL
SPACERS AS APPROPRIATE

OUTER ENCAPSULATION
TUBING .382 OD
x .025 WALL
x 16.865 LG

REVISED 3-10-95 BY JG-CHECKED BY F.S.

| | | | | | |
|---|-----------------|----------------------|---|--|--|
| UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES | | | NEUTRON PRODUCTS INC | | |
| TOLERANCES ANGULAR DECIMAL FRACTIONAL $\pm 3^\circ$.XX = $\pm .01$.XXX = $\pm .005$ | | | TITLE | | |
| DRAWN ED 11-20-90 | | | RADIATION PROCESSING SOURCE NPRP 450-10-B | | |
| DESIGN ED 11-20-90 | | | | | |
| CHECKED FS | | | | | |
| APPROVED FS | | ISSUED DATE 11/21/90 | | | |
| SIZE A | DWG. NO. 200355 | REV B | | | |
| SCALE 2X | | SHEET 1 OF 1 | | | |



U.S. Department
of Transportation

East Building, PHH-23
1200 New Jersey Avenue SE
Washington, D.C. 20590

**Pipeline and
Hazardous Materials
Safety Administration**

CERTIFICATE NUMBER: USA/0458/S-96, Revision 6

ORIGINAL REGISTRANT(S):

Mr. Edmund DeRosa
Manager, Teletherapy Operations
Neutron Products, Inc.
22301 Mt. Ephraim Road
P.O. Box 68
Dickerson, 20842
USA

Mr. Marvin Turkanis
Neutron Products, Inc.
22301 Mt. Ephraim Road
P.O. Box 68
Dickerson, 20842
USA